

日付	時刻	ANS0 (CHS-UGS)	Vref	ANS2 (LM35DZ-1)	ANS3 (LM35DZ-2)	湿度(%RH)	温度1(°C)	温度2(°C)
2010/11/13	22:26:01	471	1023	547	549	57.4	22.2	22.3
2010/11/13	22:26:02	471	1023	547	554	57.4	22.2	22.5
2010/11/13	22:26:03	471	1023	546	550	57.4	22.2	22.4
2010/11/13	22:26:04	472	1023	547	550	57.6	22.2	22.4
2010/11/13	22:26:05	471	1023	547	550	57.4	22.2	22.4
2010/11/13	22:26:06	472	1023	547	550	57.6	22.2	22.4
2010/11/13	22:26:07	472	1023	547	550	57.6	22.2	22.4
2010/11/13	22:26:08	471	1023	547	550	57.4	22.2	22.4
2010/11/13	22:26:10	471	1023	547	547	57.4	22.2	22.2
2010/11/13	22:26:11	472	1023	546	551	57.6	22.2	22.4

WORK1
 2010/11/13 23:19
 end
 ← 実行終了ENDマーク

interval time (ms)	1000
--------------------	------

読込回数 10

マクロ開始= ctrl + shift + A
 中止 = CTRL + BREAK
 マクロも変更済み
 湿度センサーはオペアンプで2倍増幅
 温度センサーはオペアンプで6倍増幅